

**Vaunix Technology Corporation**  
**Lab Brick® Family of Signal Generators**

**API User**  
**Manual**



**NOTICE**

Vaunix has prepared this manual for use by Vaunix Company personnel and customers as a guide for the customized programming of Lab Brick products. The drawings, specifications, and information contained herein are the property of Vaunix Technology Corporation, and any unauthorized use or disclosure of these drawings, specifications, and information is prohibited; they shall not be reproduced, copied, or used in whole or in part as the basis for manufacture or sale of the equipment or software programs without the prior written consent of Vaunix Technology Corporation.

# Table of Contents

<b>1.0</b>	<b>OVERVIEW .....</b>	<b>4</b>
<b>2.0</b>	<b>USING THE SDK.....</b>	<b>5</b>
<b>3.0</b>	<b>PROGRAMMING .....</b>	<b>6</b>
3.1	Overall Strategy and API Achitecture .....	6
3.2	Status Codes .....	8
3.3	Functions – Selecting the Device .....	8
3.4	Functions – Setting parameters on the Signal Generator .....	10
3.5	Functions – Reading parameters from the Signal Generator .....	12
<b>4.0</b>	<b>PROGRAMMING SUPPORT .....</b>	<b>14</b>

## 1.0 OVERVIEW

The LabBrick Signal Generator Win32 SDK supports developers who want to control LabBrick Signal Generator from Windows programs, or who want to control the synthesizers from LabVIEW<sup>1</sup> or other National Instruments programming environments. The SDK includes a dll which provides a Win32 API to find, initialize, and control the synthesizers, along with header files and an example Win32 C program which demonstrates the use of the API.

<sup>1</sup> LabView is a trademark of National Instruments

## 2.0 USING THE SDK

The SDK consists of a dll, named VNfsynth.dll, along with this documentation, a C style header file, a library file for linking to the dll, and a VC 6 example program. Unzip the SDK into a convenient place on your hard disk, and then copy the dll and library file into the directory of the executable program you are creating. Add the header file (VNX\_fsynth.h) to your project, and include it with the other header files in your program. Make sure that the linker directives include the path of the library file.

## 3.0 PROGRAMMING

### 3.1 Overall Strategy and API Achitecture

The API provides functions for identifying how many and what type of LabBrick signal generators are connected to the system, initializing signal generators so that you can send them commands and read their state, functions to control the operation of the signal generators, and finally a function to close the software connection to the signal generator when you no longer need to communicate with it.

The API can be operated in a test mode, where the functions will simulate normal operation but will not actually communicate with the hardware devices. This feature is provided as a convenience to software developers who may not have a LabBrick signal generator with them, but still want to be able to work on an applications program that uses the LabBrick. Of course it is important to make sure that the API is in its normal mode in order to access the actual hardware!

Be sure to call `fnLSG_SetTestMode(FALSE)`, unless of course you want the API to operate in its test mode. In test mode there will be 2 devices, an LSG-402 and an LSG-602.

The first step is to identify the synthesizers connected to the system. Call the function `fnLSG_GetNumDevices()` to get the number of synthesizers attached to the system. Note that USB devices can be attached and detached by users at any time. If you are writing a program which needs to handle the situation where devices are attached or detached while the program is operating, you should periodically call `fnLSG_GetNumDevices()` to see if any new devices have been attached.<sup>2</sup>

Allocate an array big enough to hold the device ids for the number of devices present. While you should use the `DEVID` type declared in `VNX_fsynth.h` it's just an array of units at this point. You may want to allocate an array large enough to hold `MAXDEVICES` device ids, so that you do not have to handle the case where the number of attached devices increases.

Call `fnLSG_GetDevInfo(DEVID *ActiveDevices)`, which will fill in the array with the device ids for each connected frequency synthesizer. The function returns an integer, which is the number of devices present on the machine.

<sup>2</sup> Usually it is a good idea to call `fnLSG_GetNumDevices()` at around 1 second intervals. While a short interval reduces the chances, it is still possible that the user will remove one device and replace it with another however, so to completely handle all the cases which can result from users hot plugging devices your application needs to check to see not only if the number of devices is different, but if the same number of devices are present, but they are different devices.

The next step is to call `fnLSG_GetModelName(DEVID deviceID, char *ModelName)` with a null `ModelName` pointer to get the length of the model name, or just use a buffer that can hold `MAX_MODELNAME` chars. You can use the model name to identify the type of synthesizer. Call `fnLSG_GetSerialNumber(DEVID deviceID)` to get the serial number of the synthesizer. Based on that information, your program can determine which device to open.

Once you have identified the synthesizer you want to send commands to, call `fnLSG_InitDevice(DEVID deviceID)` to actually open the device and get its various parameters like frequency setting, frequency step parameters, etc. After the `fnLSG_InitDevice` function has completed you can use any of the get functions to read the settings of the synthesizer.

To change one of the settings of the synthesizer, use the corresponding set function. For example, to set the synthesizer frequency, call `fnLSG_SetFrequency(DEVID deviceID, int frequency)`. The first argument is the device id of the synthesizer, the second is the desired output frequency. Frequency is specified in 100 KHz increments, where:

$$\text{frequency} = \text{Frequency (Hz)} / 100,000$$

For example, to specify an output frequency of 1.5 GHz, `frequency = 15,000`.

To set the output power level, call `fnLSG_SetPowerLevel(DEVID deviceID, int powerlevel)` with the output power level you want. The `powerlevel` is encoded as the number of .25dB increments, with a resolution of .5dB. To set a power level of +5 dBm, for example, `powerlevel` would be 20. To set a power level of -20 dBm, `powerlevel` would be -80.

Note that the LabBrick signal generators have a maximum and minimum settable power level. You can query the limits with calls to `fnLSG_GetMaxPwr(DEVID deviceID)` and `fnLSG_GetMinPwr(DEVID deviceID)`. Both functions use the same encoding of the `powerlevel` as the `SetPowerLevel` function.

When you are done with the device, call `fnLSG_CloseDevice(DEVID deviceID)`.

### 3.2 Status Codes

All of the set functions return a status code indicating whether an error occurred. The get functions normally return an integer value, but in the event of an error they will return an error code. The error codes can be distinguished from normal data by their numeric value, since all error codes have their high bit set, and they are outside of the range of normal data.

A separate function, `fnLSG_GetDeviceStatus(DEVID deviceID)` provides access to a set of status bits describing the operating state of the synthesizer. This function can be used to check if a device is currently connected or open.

The values of the status codes are defined in the `VNX_fsynth.h` header file.

### 3.3 Functions – Selecting the Device

`VNX_FSYNSTH_API void fnLSG_SetTestMode(bool testmode)`

Set testmode to FALSE for normal operation. If testmode is TRUE the dll does not communicate with the actual hardware, but simulates the basic operation of the dll functions. It does not simulate the operation of attenuation ramps generated by the actual hardware, but it does simulate the behavior of the functions used to set the parameters for the ramps.

`VNX_FSYNSTH_API int fnLSG_GetNumDevices()`

This function returns a count of the number of connected attenuators.

`VNX_FSYNSTH_API int fnLSG_GetDevInfo(DEVID *ActiveDevices)`

This function fills in the `ActiveDevices` array with the device ids for the connected synthesizers. Note that the array must be large enough to hold a device id for the number of devices returned by `fnLSG_GetNumDevices`. The function also returns the number of active devices, which can, under some circumstances, be less than the number of devices returned in the previous call to `fnLSG_GetNumDevices`.

The device ids are used to identify each device, and are used in the rest of the functions to select the device. Note that while the device ids may be small integers, and may, in some circumstances appear to be numerically related to the devices present, they should only be used as opaque handles.

VNX\_FSYNSTH\_API int fnLSG\_GetModelName(DEVID deviceID, char \*ModelName)

This function is used to get the model name of the synthesizer. If the function is called with a null pointer, it returns just the length of the model name string. If the function is called with a non-null string pointer it copies the model name into the string and returns the length of the string. The string length will never be greater than the constant MAX\_MODELNAME which is defined in VNX\_fsynth.h. This function can be used regardless of whether or not the synthesizer has been initialized with the fnLSG\_InitDevice function.

VNX\_FSYNSTH\_API int fnLSG\_GetSerialNumber(DEVID deviceID)

This function is used to get the serial number of the synthesizer. It can be called regardless of whether or not the synthesizer has been initialized with the fnLSG\_InitDevice function. If your system has multiple synthesizers, your software should use each device's serial number to keep track of each specific device. Do not rely upon the order in which the devices appear in the table of active devices. On a typical system the individual synthesizers will typically be found in the same order, but there is no guarantee that this will occur.

VNX\_FSYNSTH\_API int fnLSG\_GetDeviceStatus(DEVID deviceID)

This function can be used to obtain information about the status of a device, even before the device is initialized. (Note that information on the stepped sweep activity of the device is not guaranteed to be available before the device is initialized.)

VNX\_FSYNSTH\_API int fnLSG\_InitDevice(DEVID deviceID)

This function is used to open the device interface to the synthesizer and initialize the dll's copy of the device's settings. If the fnLSG\_InitDevice function succeeds, then you can use the various fnLSG\_Get\* functions to read the synthesizer's settings. This function will fail, and return an error code if the synthesizer has already been opened by another program.

VNX\_FSYNSTH\_API int fnLSG\_CloseDevice(DEVID deviceID)

This function closes the device interface to the synthesizer. It should be called when your program is done using the synthesizer.

### 3.4 Functions – Setting parameters on the Signal Generator

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetFrequency(DEVID deviceID, int frequency)

This function is used to set the output frequency of the synthesizer. Frequency is encoded as an integer number of 100 kHz steps:

$$\text{frequency} = \text{Frequency (Hz)} / 100,000$$

For example, to specify an output frequency of 250 MHz, frequency = 2500. The value of frequency must be within the range of the attached synthesizer or an error will be returned.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetPowerLevel(DEVID deviceID, int powerlevel);

This function is used to set the output power level of the programmable synthesizer. The power level is specified in .25dB units. The encoding is:

$$\text{powerlevel} = \text{desired output power in dBm} / .25\text{dB}$$

For example, if you want -7.5 dBm output power then you should set powerlevel to -30.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetStartFrequency(DEVID deviceID, int startfrequency)

This function sets the frequency at the beginning of a stepped frequency sweep. The encoding of startfrequency is the same as the fnLSG\_SetFrequency function. Note that the start frequency should be less than the end frequency when you want the frequency to step upwards during the sweep. For a sweep where the frequency decreases, then the start frequency should be larger than the end frequency.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetEndFrequency(DEVID deviceID, int endfrequency)

This function sets the frequency at the end of a stepped frequency sweep. The encoding of endfrequency is the same as the fnLSG\_SetFrequency function.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetFrequencyStep(DEVID deviceID, int frequencystep)

This function sets the size of the frequency step that will be used to generate the output frequency sweep. The encoding of frequencystep is the same as for frequency in the fnLSG\_SetFrequency function. The smallest step size is 1 or 100 kHz.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetDwellTime(DEVID deviceID, int dwelltime)

This function sets the length of time that the synthesizer will dwell on each frequency step while it is generating the frequency sweep. The dwelltime variable is encoded as the number of milliseconds to dwell at each frequency. The minimum dwell time is 10 milliseconds.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetRFOn(DEVID deviceID, bool on)

This function turns the RF stages of the synthesizer on (on = TRUE) or off (on = FALSE).

VNX\_FSYNSTH\_API int fnLSG\_SetUseInternalRef(DEVID deviceID, bool internal);

This function configures the synthesizer to use the internal reference if internal = 1. If internal = 0, then the synthesizer is configured to use an external frequency reference.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetSweepDirection(DEVID deviceID, bool up)

This function is used to set the direction of the stepped frequency sweep. To create a sweep with increasing frequency, set up = TRUE. Note that the sweep start frequency value must be less than the sweep end frequency value for a sweep with increasing frequency. For a sweep that decreases in frequency, the sweep start value must be greater than the sweep end value.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SetSweepMode(DEVID deviceID, bool mode)

This function is used to select either a single frequency sweep, or a repeating series of sweeps. If mode = TRUE then the sweep will be repeated, if mode = FALSE the sweep will only happen once.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_StartSweep(DEVID deviceID, bool go)

This function is used to start and stop the frequency sweeps. If go = TRUE the synthesizer will begin sweeping, FALSE stops the sweep. You must set the sweep parameters before calling this function to start the sweep.

VNX\_FSYNSTH\_API LVSTATUS fnLSG\_SaveSettings(DEVID deviceID)

The LabBrick synthesizers can save their settings, and then resume operating with the saved settings when they are powered up. Set the desired parameters, then use this function to save the settings.

### 3.5 Functions – Reading parameters from the Signal Generator

VNX\_FSYNSTH\_API int fnLSG\_GetFrequency(DEVID deviceID)

This function returns the current frequency setting of the selected device. When a sweep is active this value will change dynamically to reflect the current setting of the device. The return value is in 100 kHz units.

VNX\_FSYNSTH\_API int fnLSG\_GetStartFrequency (DEVID deviceID)

This function returns the current frequency sweep starting value setting of the selected device. The return value is in 100 kHz units.

VNX\_FSYNSTH\_API int fnLSG\_GetEndFrequency (DEVID deviceID)

This function returns the current frequency sweep end setting of the selected device. The return value is in 100 kHz units.

VNX\_FSYNSTH\_API int fnLSG\_GetFrequencyStep (DEVID deviceID)

This function returns the current frequency sweep step size setting of the selected device. The return value is in 100 kHz units, so for example a frequency step of 10 MHz would be represented by a return value of 100.

VNX\_FSYNSTH\_API int fnLSG\_GetDwellTime(DEVID deviceID)

This function returns the current dwell time for each step on the frequency ramp in milliseconds. A one second dwell time, for example, would be returned as 1000.

VNX\_FSYNSTH\_API int fnLSG\_GetRF\_On(DEVID deviceID)

This function returns an integer value which is 1 when the synthesizer is “on”, or 0 when the synthesizer has been set “off” by the fnLSG\_SetRFOn function.

VNX\_FSYNSTH\_API int fnLSG\_GetUseInternalRef(DEVID deviceID);

This function returns an integer value which is 1 when the synthesizer is configured to use its internal frequency reference. It returns a value of 0 when the synthesizer is configured to use an external frequency reference.

VNX\_FSYNSTH\_API int fnLSG\_GetPowerLevel(DEVID deviceID);

This function returns the current power level setting as an integer number of .25 dB units. As an example, an output power level of +3 dBm would result in the value 12 being returned, while an output power level of +3.5 dBm would result in the value 14 being returned. The output power resolution is .5 dB.

VNX\_FSYNSTH\_API int fnLSG\_GetMaxPwr(DEVID deviceID);

This function returns the maximum output power level that the synthesizer can provide, encoded in the same format as the fnLSG\_GetPowerLevel function. For a synthesizer with +10 dBm maximum output power level this function returns the integer value 40. This is a read only value.

VNX\_FSYNSTH\_API int fnLSG\_GetMinPwr(DEVID deviceID);

This function returns the minimum output power level that the synthesizer can provide, encoded in the same format as the fnLSG\_GetPowerLevel function. Typically this value is a negative number. For example, a device with -45 dBm minimum output power would return an integer value of -180. This is a read only value.

VNX\_FSYNSTH\_API int fnLSG\_GetMaxFreq (DEVID deviceID)

This function returns the maximum output frequency that the device can provide. The value is represented in 100 kHz units.

VNX\_FSYNSTH\_API int fnLSG\_GetMinFreq(DEVID deviceID)

This function returns the minimum output frequency that the device can provide. The value is represented in 100 kHz units.

#### **4.0 PROGRAMMING SUPPORT**

Lab Brick programming support is available from Vaunix Technology Corporation. Please contact our technical support group by email - [LabBrickSupport@Vaunix.com](mailto:LabBrickSupport@Vaunix.com).

Vaunix Technology also offers custom programming solutions. Send us your requirements to receive a fixed rate project quotation.

Thank you for using our Lab Brick products.